



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/643,261

08/20/2003

Ilya Yampolsky

BSI-540US

4665

60117

7590

06/26/2008

RATNER PRESTIA

P.O. BOX 980

VALLEY FORGE, PA 19482

EXAMINER

TRUONG, KEVIN THAO

ART UNIT

PAPER NUMBER

3734

MAIL DATE

DELIVERY MODE

06/26/2008

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/643,261	Applicant(s) YAMPOLSKY ET AL.	
	Examiner Kevin T. Truong	Art Unit 3734	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 27 February 2008.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-17 is/are pending in the application.
- 4a) Of the above claim(s) 13-17 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-12 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Note: This is in response to an amendment filed 02/27/2008.

Claim Rejections - 35 USC § 103

1. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

2. Claims 1-12 are rejected under 35 U.S.C. 103(a) as being unpatentable over Zarins et al. (US 7131991) in view of Cardon et al. (US 5383892).

Zarins et al., in figure 3B, disclose trunk region (the large diameter region which includes the top, large diameter portion of cuff 50b and the entire crown 61) having a self-expandable section (the top, large diameter portion of cuff 50b which is self-expandable as indicated in col. 7, lines 39- 40) and a section 61 extending from a first end of the self- expandable section and at least one self-expandable branch 51.

Zarins et al. is silent as to whether the crown 61 (which is used to secure the device to the blood vessel as indicated in col. 30-32 and col. 8, lines 18-20) is balloon expandable or not. However, Cardon et al. teach that the end portion of a hybrid stent device (a stent device that includes both self- expandable and balloon expandable portions) should be balloon expandable in order to obtain the advantage of insuring that the device is anchored to the blood vessel (col. 2, lines 54-59). It would have been obvious to make the crown portion 61 of the Zarins et al. stent device balloon expandable so that it too would have this advantage. With this modification, the Zarins et al. trunk

region of the stent device would include a balloon expandable section as claimed in claim i, line 3-4. As to claim 6, Zarins et al. fail to disclose a crimping member. However, it is old and well known in this art to use crimping members to secure prosthesis sections together since they provide a strong attachment. It would have been obvious to include a crimping member in the Zarins et al. stent so that it too would have this advantage. The above well known in the art statement is taken to be admitted prior art because applicant failed to traverse the examiner's assertion (M.P.E.P. 2144.03). As to claim 11, note col. 2, lines 60 to col. 3, line 18 of Cardon et al.

Response to Arguments

3. Applicant's arguments filed 02/27/2008 have been fully considered but they are not persuasive. In response to applicant's argument that there is no suggestion to combine the references, the examiner recognizes that obviousness can only be established by combining or modifying the teachings of the prior art to produce the claimed invention where there is some teaching, suggestion, or motivation to do so found either in the references themselves or in the knowledge generally available to one of ordinary skill in the art. See *In re Fine*, 837 F.2d 1071, 5 USPQ2d 1596 (Fed. Cir. 1988) and *In re Jones*, 958 F.2d 347, 21 USPQ2d 1941 (Fed. Cir. 1992). In this case, the combination of Zarins and Cardon is considered providing a prima facie case of obviousness as stated in the above rejection.

4. In response to applicant's argument that the present invention discloses the benefit of the self-expandable sections being to allow the self-expandable sections to conform to the shape of the body lumens surrounding them at their deployment site and

to easily advance through the tortuous confines of body lumens. On the other hand, Cardon discloses the rigid ends help the stent to be positioned and ensure its anchoring which teaches away from the present claim 1, the fact that applicant has recognized another advantage which would flow naturally from following the suggestion of the prior art cannot be the basis for patentability when the differences would otherwise be obvious. See *Ex parte Obiaya*, 227 USPQ 58, 60 (Bd. Pat. App. & Inter. 1985).

Conclusion

5. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Kevin T. Truong whose telephone number is 571-272-4705. The examiner can normally be reached on Monday-Friday from 8:00 AM to 4:30 PM..

The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Kevin T. Truong/
Primary Examiner, Art Unit 3734

Kevin T. Truong
Primary Examiner
Art Unit 3734